

## EON PFAS-Free De-Ionized Water Lab Results

**Client Sample ID: 5M7-56**  
**Lab Sample ID: 320-48897-5**  
**Matrix: Water**

**eurofins-TestAmerica**

<b>Method: 537 (modified) - Fluorinated Alkyl Substances</b>					
Analyte	Result	Qualifier	RL	MDL	Unit
Perfluorobutanoic acid (PFBA)	ND		1.8	0.32	ng/L
Perfluorooctanoic acid (PFOA)	ND		1.8	0.77	ng/L
Perfluorooctane sulfonate (PFOS)	ND		1.8	0.49	ng/L
Perfluoropentanoic acid (PFPeA)	ND		1.8	0.44	ng/L
Perfluorohexanoic acid (PFHxA)	ND		1.8	0.52	ng/L
Perfluoroheptanoic acid (PFHpA)	ND		1.8	0.23	ng/L
Perfluorobutanesulfonic acid (PFBS)	ND		1.8	0.18	ng/L
Perfluorohexanesulfonic acid (PFHxS)	0.25	J B	1.8	0.15	ng/L
6:2 FTS	ND		18	1.8	ng/L
Perfluoroundecanoic acid (PFUnA)	ND		1.8	0.99	ng/L
Perfluorodecanesulfonic acid (PFDS)	ND		1.8	0.29	ng/L
Perfluorotetradecanoic acid (PFTeA)	ND		1.8	0.26	ng/L
Perfluorodecanoic acid (PFDA)	ND		1.8	0.28	ng/L
Perfluorotridecanoic acid (PFTriA)	ND		1.8	1.2	ng/L
4:2 FTS	ND		18	4.7	ng/L
Perfluorononanoic acid (PFNA)	ND		1.8	0.24	ng/L
Perfluoropentanesulfonic acid (PFPeS)	ND		1.8	0.27	ng/L
Perfluorononanesulfonic acid (PFNS)	ND		1.8	0.14	ng/L
8:2 FTS	ND		18	1.8	ng/L
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	ND		18	1.7	ng/L
Perfluorododecanoic acid (PFDoA)	ND		1.8	0.50	ng/L
Perfluorooctanesulfonamide (FOSA)	ND		1.8	0.32	ng/L
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	ND		18	2.8	ng/L
Perfluoroheptanesulfonic Acid (PFHpS)	ND		1.8	0.17	ng/L

Qualifier	Qualifier Description				
<b>B</b>	Compound was found in the blank and sample.				
<b>I</b>	Value is EMPC (estimated maximum possible concentration).				
<b>J</b>	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.				
QC Method Blank					
QC Method Blank	Result	Qualifier	RL	MDL	Unit
Perfluorohexanesulfonic acid (PFHxS)	0.357	J	2.0	0.17	ng/L