

## Equilibrator™ PDB PFAS-Free Test Results

**Client Sample ID: 5M7-76**  
**Lab Sample ID: 320-49606-12**  
**Matrix: Water**

**eurofins-TestAmerica**

**Method: 537 (modified) - Fluorinated Alkyl Substances**

Analyte	Result	Qualifier	RL	MDL	Unit
Perfluorobutanoic acid (PFBA)	0.55	J B	1.8	0.32	ng/L
Perfluoropentanoic acid (PFPeA)	ND		1.8	0.45	ng/L
Perfluorohexanoic acid (PFHxA)	ND		1.8	0.53	ng/L
Perfluoroheptanoic acid (PFHpA)	ND		1.8	0.23	ng/L
Perfluorooctanoic acid (PFOA)	ND		1.8	0.78	ng/L
Perfluorononanoic acid (PFNA)	ND		1.8	0.25	ng/L
Perfluorodecanoic acid (PFDA)	ND		1.8	0.28	ng/L
Perfluoroundecanoic acid (PFUnA)	ND		1.8	1.0	ng/L
Perfluorododecanoic acid (PFDoA)	ND		1.8	0.51	ng/L
Perfluorotridecanoic acid (PFTriA)	ND		1.8	1.2	ng/L
Perfluorotetradecanoic acid (PFTeA)	ND		1.8	0.27	ng/L
Perfluorobutanesulfonic acid (PFBS)	ND		1.8	0.18	ng/L
Perfluoropentanesulfonic acid (PFPeS)	ND		1.8	0.28	ng/L
Perfluorohexanesulfonic acid (PFHxS)	0.30	J B	1.8	0.16	ng/L
Perfluoroheptanesulfonic Acid (PFHpS)	ND		1.8	0.17	ng/L
Perfluorooctanesulfonic acid (PFOS)	ND		1.8	0.50	ng/L
Perfluorononanesulfonic acid (PFNS)	ND		1.8	0.15	ng/L
Perfluorodecanesulfonic acid (PFDS)	ND		1.8	0.29	ng/L
Perfluorooctanesulfonamide (FOSA)	ND		1.8	0.32	ng/L
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	ND		18	2.8	ng/L
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	ND		18	1.7	ng/L
4:2 FTS	ND		18	4.8	ng/L
6:2 FTS	ND		18	1.8	ng/L
8:2 FTS	ND		18	1.8	ng/L

Qualifier	Qualifier Description
<b>B</b>	Compound was found in the blank and sample.
<b>I</b>	Value is EMPC (estimated maximum possible concentration).
<b>J</b>	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

  

QC Method Blank	Result	Qualifier	RL	MDL	Unit
Perfluorobutanoic acid (PFBA)	0.476	J	2.0	0.35	ng/L
Perfluorohexanesulfonic acid (PFHxS)	0.357	J	2.0	0.17	ng/L