

Rinsate Test Results PFAS-Free Dual Membrane Diffusion Sampler (HDPE)

Client Sample ID: 5M7-36
Lab Sample ID: 320-48897-3
Matrix: Water

eurofins-TestAmerica

Method: 537 (modified) - Fluorinated Alkyl Substances

Analyte	Result	Qualifier	RL	MDL	Unit
Perfluorobutanoic acid (PFBA)	ND		1.7	0.30	ng/L
Perfluorooctanoic acid (PFOA)	ND		1.7	0.72	ng/L
Perfluorooctane sulfonate (PFOS)	ND		1.7	0.46	ng/L
Perfluoropentanoic acid (PFPeA)	ND		1.7	0.42	ng/L
Perfluorohexanoic acid (PFHxA)	ND		1.7	0.49	ng/L
Perfluoroheptanoic acid (PFHpA)	ND		1.7	0.21	ng/L
Perfluorobutanesulfonic acid (PFBS)	ND		1.7	0.17	ng/L
Perfluorohexanesulfonic acid (PFHxS)	0.27	J B	1.7	0.14	ng/L
6:2 FTS	ND		17	1.7	ng/L
Perfluoroundecanoic acid (PFUnA)	ND		1.7	0.94	ng/L
Perfluorodecanesulfonic acid (PFDS)	ND		1.7	0.27	ng/L
Perfluorotetradecanoic acid (PFTeA)	ND		1.7	0.25	ng/L
Perfluorodecanoic acid (PFDA)	ND		1.7	0.26	ng/L
Perfluorotridecanoic acid (PFTriA)	ND		1.7	1.1	ng/L
4:2 FTS	ND		17	4.4	ng/L
Perfluorononanoic acid (PFNA)	ND		1.7	0.23	ng/L
Perfluoropentanesulfonic acid (PFPeS)	ND		1.7	0.26	ng/L
Perfluorononanesulfonic acid (PFNS)	ND		1.7	0.14	ng/L
8:2 FTS	ND		17	1.7	ng/L
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	ND		17	1.6	ng/L
Perfluorododecanoic acid (PFDoA)	ND		1.7	0.47	ng/L
Perfluorooctanesulfonamide (FOSA)	ND		1.7	0.30	ng/L
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	ND		17	2.6	ng/L
Perfluoroheptanesulfonic Acid	ND		1.7	0.16	ng/L

Qualifier	Qualifier Description
B	Compound was found in the blank and sample.
I	Value is EMPC (estimated maximum possible concentration).
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

QC Method Blank	Result	Qualifier	RL	MDL	Unit
Perfluorohexanesulfonic acid (PFHxS)	0.357	2	0.2	0.17	ng/L